Set Configuration: Metric



Set Configuration: Inch



Specifications

М	17	7	r

Code No.	Graduation	Range	Accuracy	Scale	Measuring force		
Vertical type, basic set							
513-454-10E	0.01 mm	0.8 mm	8 µm	0-40-0	0.3 N or less		
513-455-10E	0.002 mm	0.2 mm	3 µm	0-100-0	0.5 N OF IESS		
Vertical type, full set							
513-454-10T	0.01 mm	0.8 mm	8 µm	0-40-0	- 0.3 N or less		
513-455-10T	0.002 mm	0.2 mm	3 µm	0-100-0	- 0.3 IN OF IESS		
Parallel type, basic set							
513-484-10E	0.01 mm	0.8 mm	8 µm	0-40-0	0.3 N or less		
Parallel type, full set							
513-484-10T	0.01 mm	0.8 mm	8 µm	0-40-0	0.3 N or less		
Inch							
Code No	Graduation	Range	Accuracy	Scalo	Measuring force		

Code No.	Graduation	Range	Accuracy	Scale	Measuring force		
Vertical type, basic set							
513-452-10E	.0005"	.03"	±.0005"	0-15-0	0.3 N or less		
513-453-10E	.0001 "	.008"	±.0001"	0-4-0	U.3 IN OI less		
Vertical type, full set							
513-452-10T	.0005"	.03"	±.0005"	0-15-0	- 0.3 N or less		
513-453-10T	.0001"	.008"	±.0001"	0-4-0	0.514 01 less		
Parallel type, full set							
513-482-10T	.0005"	.03"	±.0005"	0-15-0	0.3 N or less		

